



# Standard Test Method for Acidity-Alkalinity of Halogenated Organic Solvents and Their Admixtures<sup>1</sup>

This standard is issued under the fixed designation D 2989; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

*This standard has been approved for use by agencies of the Department of Defense.*

<sup>ε1</sup> NOTE—Editorial corrections were made in March 2002.

## 1. Scope

1.1 This test method covers the determination of acidity in halogenated organic solvents and admixtures thereof. The alkalinity may be determined utilizing Test Method D 2106, by substituting the end point measured at pH 7 by bromothymol blue or pH meter.

1.2 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* Specific precautionary statements are given in Section 7.

## 2. Referenced Documents

### 2.1 ASTM Standards:

- D 2106 Test Method for the Determination of Amine Acid Acceptance (Alkalinity) of Halogenated Organic Solvents<sup>2</sup>
- D 2110 Test Method for pH of Water Extractions of Halogenated Organic Solvents and Their Admixtures<sup>2</sup>

## 3. Summary of Test Method

3.1 A sample of halogenated solvent or admixture is measured for pH using Test Method D 2110. If the pH of the sample is above 7.0, the alkalinity is determined using Test Method D 2106 (to an end point of pH 7). If the pH is below 7.0, the free acid content of the halogenated organic solvent or admixture is determined after water extraction using Procedure A, or can be determined directly using Procedure B.

3.1.1 *Procedure A*, using glass electrode pH meter, or

3.1.2 *Procedure B*, anhydrous methanolic sodium hydroxide titration.

## 4. Significance and Use

4.1 This test method can be used to establish manufacturing

and purchasing specifications. It can also be used to determine the condition of solvents in use.

## 5. Apparatus

- 5.1 *Separatory Funnel*, 250-mL.
- 5.2 *Graduated Cylinder*, 100 mL.
- 5.3 *Volumetric Pipets*, 1 mL, 10 mL, 25 mL, 50 mL.
- 5.4 *Beaker*, 100 mL.
- 5.5 *Borosilicate or Stainless Steel Beaker*, 2 L.
- 5.6 *Erlenmeyer Flask*, 100 mL.
- 5.7 *pH Meter with pH Electrodes*.
- 5.8 *Buret*, 10 mL.
- 5.9 *Volumetric Flask*, 100 mL, 1 L.
- 5.10 *Micro Buret*, 5 mL, Class A or *Syringe*, 100  $\mu$ L.

## 6. Reagents

6.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.<sup>3</sup> Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

6.2 *Water*—Prepare neutral, distilled or deionized water as follows: Boil 1 L of distilled or deionized water for 5 min in a borosilicate glass or stainless steel container, then cover and cool to room temperature. Titrate to a pH of 7.0 to 7.3 with either 0.01 *N* sodium hydroxide (NaOH) solution or 0.01 *N* hydrochloric acid (HCl), depending upon whether the initial pH is above or below 7.

6.3 *Bromothymol Blue Indicator Solution (0.1 %)*—Indicator solution may be purchased from a laboratory supply house or prepared as follows: Dissolve 0.1 g of dibromothymol

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D26 on Halogenated Organic Solvents and Fire Extinguishing Agents and is the direct responsibility of Subcommittee D26.04 on Test Methods.

Current edition approved Dec. 10, 2001. Published February 2002. Originally published as D 2989 – 71 T. Last previous edition D 2989 – 97.

<sup>2</sup> *Annual Book of ASTM Standards*, Vol 15.05.

<sup>3</sup> *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

sulfonephthalein in 1.6 mL of 0.1 N NaOH solution and dilute to 100 mL with methanol.

6.4 *Sodium Hydroxide Solution (0.01 N), Procedure A*—Dissolve 4 g of sodium hydroxide (NaOH) in boiled, neutral, distilled or deionized water and dilute to 1000 mL. Prepare 0.01 N NaOH solution by diluting 10.0 mL of the 0.1 N NaOH to 100 mL with boiled, neutral, distilled or deionized water in a calibrated volumetric flask. Standardize this 0.01 N NaOH by any accepted procedure to determine the exact normality. This reagent is also available from most supply houses. Protect the reagent from absorption of atmospheric carbon dioxide and standardize weekly.

6.5 *Sodium Hydroxide Methanolic Solution (0.01 N), Procedure B*—Dissolve 4 g of ACS reagent grade sodium hydroxide (NaOH) in ACS reagent grade anhydrous methanol and dilute to 1000 mL with ACS reagent grade anhydrous methanol. Prepare 0.01 N methanolic NaOH solution by diluting 10.0 mL of the 0.1 N NaOH to 100 mL with ACS reagent grade anhydrous methanol in a calibrated volumetric flask. Standardize this 0.01 N NaOH by any accepted procedure to determine the exact normality. This reagent is also available from most supply houses. Protect the reagent from absorption of atmospheric carbon dioxide and standardize weekly.

## 7. Hazards

7.1 *Solvent Hazards*—Contact with skin should be avoided to prevent removal of natural oils. Solvents are not flammable, but each has a threshold limit value for contact with vapors. These threshold limits should be determined before a solvent is listed.

7.2 NaOH as a solid or in solution can be hazardous if there is skin contact.

## 8. Procedure

8.1 *Procedure A*—Pipet 50 mL of halogenated organic solvent or admixture into a 250 mL separatory funnel containing 50 mL of freshly boiled, neutral (pH 7.0 to 7.3), distilled or deionized water. Shake the mixture for 2 min, then let stand and allow the layers to separate. Drain and discard the lower, organic layer from the separatory funnel and transfer the water layer to a 100-mL beaker. Titrate the sample, with stirring, with 0.01 N NaOH using a 5 mL Class A buret until the pH is stable between 7.0 and 7.3 for 30 s. A 100 µL syringe can be substituted for the micro buret. This technique is often used to

measure very low levels of acidity, below 1 ppm. Technique is very important. Titrate slowly, making sure the titrant falls cleanly into the water being titrated. Record the titer and calculate the acidity in accordance with 9.1.

8.2 *Procedure B*—Pipet 50 mL of halogenated organic solvent or admixture into a clean, dry, 100-mL Erlenmeyer flask. Add 1 mL of bromothymol blue indicator solution and titrate, with stirring, with 0.01 N methanolic NaOH solution to the bromothymol blue end point (see Note 1).

NOTE 1—The bromothymol blue end point may be interpreted as an aqua (blue-green) to light blue color.

## 9. Calculation

9.1 *Procedure*—Calculate the acidity of halogenated organic solvents or admixtures as equivalent hydrochloric acid in weight percent or weight ppm as follows:

$$\text{equivalent HCl, weight \%} = [V \times N \times 0.0365] / W \times 100 \quad (1)$$

$$\text{equivalent HCl, ppm} = [V \times N \times 0.0365] / W \times 1\,000\,000$$

where:

$V$  = Volume in mL, NaOH solution required for titration of the sample,

$N$  = Normality of the NaOH solution, and

$W$  = Weight of the sample used, in g as weighed or  $g$  = sample volume in mL times its density in g per mL.

## 10. Precision <sup>4</sup>

10.1 Using Method A, one standard deviation between results of a sample run by different laboratories should be no more than about ±0.5 ppm. Within a laboratory, one standard deviation between results should be no more than about 0.2 ppm. Bias was not determined in this study.

10.2 Using Method B, one standard deviation between results of a sample run by different laboratories should be no more than about ±1.0 ppm. Within a laboratory, one standard deviation between results should be no more than about 0.1 ppm. Bias for this method is approximately +1 ppm.

## 11. Keywords

11.1 acidity; alkalinity; bromothymol blue; halogenated solvents; hydrochloric acid; pH meter; sodium hydroxide; titration

<sup>4</sup> A research report is available from ASTM headquarters. Request RR: D26-1016.

*ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.*

*This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.*

*This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).*