



Standard Test Method for Total Acid Number of Trichlorotrifluoroethane¹

This standard is issued under the fixed designation D 3444; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the determination of total acidity in trichlorotrifluoroethane and other halocarbons that are liquid at room temperature.

1.2 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* Specific precautionary statements are given in Section 6.

2. Summary of Test Method

2.1 The total acidity of the solvent is determined by titration with standard alcoholic potassium hydroxide solution using phenolphthalein solution as the indicator.

3. Significance and Use

3.1 This test method determines the total acidity of trichlorotrifluoroethane by titration with alcoholic potassium hydroxide using phenolphthalein for an indicator.

3.2 This test method can be used to develop production and purchasing specifications.

4. Apparatus

4.1 *Buret*, 2-mL in 0.01-mL subdivisions, self-leveling, with 1000-mL reservoir.

4.2 *Vinyl Tubing*.²

4.3 *Two Indicating Carbon Dioxide Absorbent Tubes*, 30–50 mesh.

5. Reagents

5.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society,

where such specifications are available.³ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

5.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean halide-free distilled water.

5.3 *Isopropyl Alcohol*.

5.4 *Phenolphthalein Indicator Solution* (0.05 g in 100 mL of isopropyl alcohol).

5.5 *Potassium Acid Phthalate*, U.S. National Bureau of Standards primary standard. Dry this chemical in a 120°C oven for several hours prior to use. Store the dry chemical in a desiccator.

5.6 *Potassium Hydroxide Pellets*.

5.7 *Potassium Hydroxide Solution in Isopropyl Alcohol* (0.01 N)—Add approximately 1 g of solid KOH to approximately 1 L of anhydrous isopropyl alcohol in a 2-L Erlenmeyer flask. Boil the mixture gently for 5 to 10 min, stirring to prevent the solids from forming a cake on the bottom. Add at least 2 g of barium hydroxide [Ba(OH)₂] and boil again gently for 5 to 10 min. Cool to room temperature. Allow to stand for several hours and filter the supernatant liquid through a fine sintered glass or porcelain filtering funnel. Avoid unnecessary exposure to CO₂ during filtration. Store the solution out of contact with cork or rubber. Protect the solution from CO₂ by a guard tube containing Ascarite II.⁴ Standardize against pure potassium acid phthalate in about 25 mL of CO₂-free water using phenolphthalein to detect the end point.

6. Hazards

6.1 Trichlorotrifluoroethane removes natural oils from skin. Avoid skin contact. The solvent is not flammable and has a threshold limit value of 1000 ppm by volume in air. See the manufacturer's MSDS.

6.2 Isopropyl alcohol is flammable, and has a threshold limit value of 400 ppm by volume in air. See the manufacturer's MSDS.

6.3 Potassium hydroxide is corrosive to body tissue. Avoid

¹ This test method is under the jurisdiction of ASTM Committee D-26 on Halogenated Organic Solvents and Fire Extinguishing Agents and is the direct responsibility of Subcommittee D26.04 on Test Methods.

Current edition approved Jan. 10, 2000. Published April 2000. Originally published as D 3444 – 75 T. Last previous edition D 3444 – 91 (1995) ^{ϵ 1}.

² Tygon has been found satisfactory. An equivalent may be used.

³ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

⁴ Ascarite II has been found satisfactory and is available from chemical supply houses.

contact with the solid and its solution. Should this chemical contact the person, immediately flush the affected area with copious amounts of water. See the manufacturer's MSDS.

7. Procedure

7.1 For sample analysis, add 50 mL of isopropyl alcohol to a 250-ml Erlenmeyer flask, add 5 drops of phenolphthalein indicator solution and titrate to a pink end point with alcoholic KOH solution.

7.1.1 Swirl flask contents gently during titration. The end point should last for 10 to 15 s. Stopper the flask until the sample is added to minimize CO₂ contamination.

7.2 Pour a weighed sample (about 50.00 ± 0.02 g) into the flask containing the pretitrated alcohol.

7.3 Titrate the sample with the standard alcoholic KOH solution to a 15-s pink end point. Record the amount of titrant required.

7.4 For standardization of 0.01 N KOH in isopropyl alcohol, proceed as follows:

7.4.1 Weigh (±0.0001 g) 0.142 ± 0.005 g of dry (National Bureau of Standards) primary standard potassium acid phthalate into a 1-L volumetric flask containing about 500 mL of deionized CO₂-free water, then stopper the flask.

7.4.2 Swirl the flask to dissolve the potassium acid phthalate.

7.4.3 Fill the flask to volume with deionized, CO₂-free water, stopper the flask then mix the solution thoroughly.

7.4.4 Place 25 mL (transfer pipet) of deionized, CO₂-free water and 5 drops of phenolphthalein indicator solution into a 250-mL Erlenmeyer flask.

7.4.5 Pipet 25 mL (transfer pipet) of the potassium acid phthalate standard solution into the flask, then seal the flask with a stopper.⁵

7.4.6 Treat the blank and standard alike from this point on.

7.4.7 Titrate the standard with the approximate 0.01 N potassium hydroxide solution to the phenolphthalein indicator end point. Record the titer.

8. Calculation

8.1 Calculate the acid number of the solvent as follows:

$$\begin{aligned} \text{Total acid number in 1 mg of KOH/g sample} \\ = (A \times N \times 56.1)/W \end{aligned}$$

where:

A = millilitres of KOH solution required for titration of sample, mL,

N = normality of KOH solution, and

W = weight of sample, g.

8.2 Calculate the normality of the potassium hydroxide standard as follows:

$$N, \text{ KOH} = (G \times V) / [204.224 \times (A - B)] \quad (1)$$

where:

A = standard titer, mL,

B = blank titer, mL,

G = potassium acid phthalate, g,

V = potassium acid phthalate aliquot, mL, and

204.224 = potassium acid phthalate/mole, g.

8.3 Average the normality values obtained from triplicate analyses to the nearest 0.0001 unit. Use the average normality for sample calculation.

9. Precision and Bias ⁶

9.1 *Repeatability (Single Analyst)*: At a mean value of 0.0018 mg-KOH/g the standard deviation of results (each the average of triplicates obtained by the same analyst on two different days) has been estimated to be 0.0010 mg-KOH/g at three degrees of freedom. Two such values should be considered suspect (95 % confidence level) if they differ by more than 0.0045 mg-KOH/g.

9.2 *Reproducibility (Multilaboratory)*: At a mean value of 0.0018 mg-KOH/g the standard deviation of results (each the average of triplicates in three different laboratories), obtained by analysts in different laboratories, has been estimated to be 0.0036 mg-KOH/g at two degrees of freedom. Two such values should be considered suspect (95 % confidence level) if they differ by more than 0.0219 mg-KOH/g.

10. Keywords

10.1 acidity; CFC-113; total acid number; trichlorotrifluoroethane

⁵ Neoprene has been found satisfactory. An equivalent may be used.

⁶ Supporting data are available from ASTM Headquarters. Request RR: D26-1011.

The American Society for Testing and Materials takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).