



Standard Test Method for Base Number Determination by Potentiometric Titration¹

This standard is issued under the fixed designation D 4739; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method covers a procedure for the determination of basic constituents in petroleum products and lubricants (Note 1). The test method resolves these constituents into groups having weak-base and strong-base ionization properties, provided the dissociation constants of the more strongly basic compounds are at least 1000 times that of the next weaker groups.

1.1.1 This test method covers base numbers up to 70. While it can be extended to higher base numbers, the precision of the test method for base numbers greater than 70 has not been determined.

NOTE 1—In new and used oils, the constituents which can be considered to have basic properties are primarily organic and inorganic bases, including amino compounds, although certain salts of heavy metals, salts of weak acids, basic salts of polyacidic compounds, and some additives such as inhibitors or detergents may show basic characteristics.

1.2 This test method can be used to indicate relative changes that occur in an oil during use under oxidizing or other service conditions regardless of the color or other properties of the resulting oil (Note 3). Although the analysis is made under closely specified conditions, the method is not intended to, and does not, result in reported basic properties which can be used under all service conditions to predict performance of an oil; for example, no overall relationship is known between bearing corrosion or the control of corrosive wear in the engine and base number.

NOTE 2—Test Method D 4739 was developed as an alternative for the former base number portion of Test Method D 664. Base numbers obtained by this method may or may not be numerically the same as those obtained by the former base number portion of Test Method D 664.²

NOTE 3—A color indicator titration method is also available in the Test Method D 974 and IP 139. The base numbers obtained by the potentiometric method may or may not be numerically the same as those obtained by Test Method D 974 or equivalent color indicator methods such as given in Federal Test Method Std. No. 791b. Potentiometric methods for base number are also available in Test Method D 2896.

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products and Lubricants and is the direct responsibility of Subcommittee D02.06 on Analysis of Lubricants.

Current edition approved Dec. 10, 2002. Published February 2003. Originally approved in 1987. Last previous edition approved in 1996 as D 4739–96.

² The base number portion was last published in the 1981 version.

1.3 The values stated in SI units are to be regarded as the standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:

D 664 Test Method for Acid Number of Petroleum Products by Potentiometric Titration³

D 974 Test Method for Acid and Base Number by Color-Indicator Titration³

D 1193 Specification for Reagent Water⁴

D 2896 Test Method for Base Number of Petroleum Products by Potentiometric Perchloric Acid Titration³

2.2 IP Standard:

IP 139 Test Method for Acid Number by Color-Indicator Titration Method⁵

2.3 U.S. Federal Test Method:

Federal Test Method Standard No. 791b Lubricants Liquid Fuels and Related Products; Methods of Testing⁶

3. Terminology

3.1 Definitions:

3.1.1 *base numbers, n*—the quantity of acid, expressed in milligrams of potassium hydroxide per gram of sample that is required to titrate a sample, dissolved in a specified solvent to a specified end point.

3.1.1.1 *Discussion*—In this test method, the sample is titrated to a meter reading corresponding to a freshly prepared nonaqueous acidic buffer solution.

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *strong base number, n*—the quantity of acid, expressed in terms of the equivalent number of milligrams of

³ *Annual Book of ASTM Standards*, Vol 05.01.

⁴ *Annual Book of ASTM Standards*, Vol 11.01.

⁵ IP Standards for Petroleum and Its Products, Part 1. Methods for Analysis and Testing.

⁶ Available from Standardization Documents Order Desk, DODSSP, Bldg. 4, Section D, 700 Robbins Ave., Philadelphia, PA 19111-5098

*A Summary of Changes section appears at the end of this standard.

potassium hydroxide per gram of sample, that is required to titrate a sample dissolved in the specified solvent from the initial meter reading to a meter reading corresponding to a freshly prepared basic buffer solution.

4. Summary of Test Method

4.1 The sample is dissolved in a mixture of toluene, propanol-2-ol (isopropyl alcohol), chloroform, and a small amount of water and titrated potentiometrically with alcoholic hydrochloric acid solution. The test results of this procedure are obtained by titration mode of fixed increment and fixed time additions of the titrant. An endpoint is selected from a titration curve according to the criteria given in 13.1 and used to calculate a base number.

5. Significance and Use

5.1 New and used petroleum products can contain basic constituents that are present as additives or as degradation products formed during service. The relative amount of these materials can be determined by titrating with acids. The base number is a measure of the amount of basic substances in the oil—always under the conditions of the test. The base number is used as a guide in the quality control of lubricating oil formulations. It is also sometimes used as a measure of lubricant degradation in service. Any condemning limits must be empirically established.

6. Apparatus

6.1 *Potentiometric Titrimeter*, automatic or manual, with capability of adding fixed increments of titrant at fixed time intervals (see Annex A1).

6.1.1 The titrimer must automatically (or manually) control the rate of addition of titrant as follows: Delivery of titrant will be incremental; after delivery of precisely a 0.100-mL increment (see 6.1.2), the delivery is stopped and a fixed time period of 90 s is allowed to pass before another 0.100-mL increment of titrant is delivered. This procedure is repeated until the titration is completed.

6.1.2 The precision of addition of the 0.100-mL increments of titrant must be ± 0.001 mL for automatic titrators. For manual buret, it should be ± 0.005 mL. A higher incremental precision is required for an automatic buret because the total volume to the end point is summed from the individual increments, whereas with a manual buret it is read from a scale.

6.2 *Glass Indicating Electrode*, pH 0 to 14, general purpose.

6.3 *Reference Electrode*, Silver/Silver Chloride (Ag/AgCl) reference electrode, filled with 1 M - 3 M LiCl in ethanol.

NOTE 4—Certain alternative electrode-electrolyte combinations have been found to give satisfactory results although the precision using these alternatives has not been determined. Combination electrodes may be used provided they conform to 8.3 and have a sufficient fast response time.

6.4 *Stirrer, Buret, Stand, Titration Vessel*, as specified in Annex A1 are required. A typical cell assembly is shown in Fig. 1.

7. Reagents

7.1 *Buffer, Nonaqueous Acid*—Add 10 mL of buffer stock solution A (see 7.3) to 100 mL of titration solvent. Use within 1 h.




FIG. 1 Cell for Potentiometric Titration

7.2 *Buffer, Nonaqueous Base*—Add 10 mL of buffer stock solution B (see 7.4) to 100 mL of titration solvent. Use within 1 h.

7.3 *Buffer Stock Solution A*—Accurately weigh 24.2 ± 0.1 g of 2,4,6-trimethyl pyridine (γ -collidine), and transfer to a 1-L volumetric flask containing 100 mL of anhydrous isopropyl alcohol. Using a 250-mL graduated cylinder, add to the flask, while continuously stirring its contents, $150/N_{\text{HCl}} \pm 5$ mL of 0.2 M alcoholic HCl solution (N_{HCl} being the exact molarity of the HCl solution found by standardization). Dilute to the 1000-mL mark with anhydrous isopropyl alcohol, and mix thoroughly. Use within 2 weeks.

7.4 *Buffer Stock Solution B*—Accurately weigh 27.8 ± 0.1 g of *m*-nitrophenol and transfer to a 1-L volumetric flask containing 100 mL of anhydrous isopropyl alcohol. Using a 250-mL graduated cylinder, add to the flask while continuously stirring its contents, $50/N_{\text{KOH}} \pm 1$ mL of 0.2 M alcoholic KOH solution. (N_{KOH} being the exact molarity of the KOH solution found by standardization). Dilute to the 1000-mL mark with anhydrous isopropyl alcohol and mix thoroughly. Use within 2 weeks.

7.5 *Chloroform*, reagent grade. (**Warning**—Toxic and suspected carcinogen.)

7.6 *Chromic Acid Solution*—(**Warning**—Causes severe burns. Recognized carcinogen. Strong oxidizer.)

7.7 *Hydrochloric Acid Solution, Standard Alcoholic (0.1 M)*—Mix 9 mL of reagent grade hydrochloric acid (HCl, sp gr 1.19) (**Warning**—Toxic and corrosive), with 1 L of anhydrous isopropyl alcohol. Standardize frequently enough to detect normality changes of 0.0005 by potentiometric titration of approximately 8 mL (accurately measured) of the 0.1 M alcoholic KOH solution diluted with 125 mL CO₂-free water.

7.8 *Hydrochloric Acid Solution, Standard Alcoholic (0.2 M)*—Prepare and standardize as described in 7.7, but use 18 mL of HCl (sp gr 1.19) .

7.9 *Lithium Chloride Electrolyte*—Prepare a saturated solution of lithium chloride (LiCl) in isopropyl alcohol.

7.10 *m*-Nitrophenol, $\text{NO}_2\text{C}_6\text{H}_4\text{OH}$, (**Warning**—Toxic), (mol weight 139.11), conforming to the following requirements:

Melting point	96 to 97°C
Color	Pale Yellow

7.10.1 Store the reagent in a brown glass bottle.

7.11 *Potassium Hydroxide Solution, Standard Alcoholic* (0.1 M)—Add 6 g of reagent grade potassium hydroxide (KOH) (**Warning**—Toxic and corrosive), to approximately 1 L of anhydrous isopropyl alcohol. Boil gently for 10 min to effect solution. Allow the solution to stand for 2 days and then filter the supernatant liquid through a fine sintered-glass funnel. Store the solution in a chemically resistant bottle. Dispense in a manner such that the solution is protected from atmospheric carbon dioxide (CO_2) by means of a guard tube containing soda lime or soda non-fibrous silicate absorbent (Ascarite, Carbosorb, or Indicarb), and such that it does not come into contact with cork, rubber, or saponifiable stopcock grease. Standardize frequently enough to detect normality changes of 0.0005 by potentiometric titration of weighed quantities of potassium acid phthalate dissolved in CO_2 -free water.

7.12 *Potassium Hydroxide Solution, Standard Alcoholic* (0.2 M)—Prepare, store, and standardize as directed in 7.11, but use 12 to 13 g of KOH to approximately 1 L of anhydrous isopropyl alcohol.

7.13 *Propanol-2-ol (Isopropyl Alcohol), Anhydrous*, (less than 0.1 % H_2O) (**Warning**—Flammable). If dry reagent cannot be procured, dry it by distillation through a multiple plate column, discarding the first 5 % of material distilling over and using the 95 % remaining. Also, drying can be accomplished using molecular sieves such as Linde Type 4A, by passing the solvent upward through a molecular sieve column using 1 part of molecular sieve per 10 parts of solvent. (**Warning**—It has been reported that, if not inhibited against it, propanol-2-ol can contain peroxides. When this occurs, an explosive mixture is possible when the storage vessel or other equipment such as a dispensing bottle, are near empty and approaching dryness.)

7.14 *Purity of Reagents*—Reagent-grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁷ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

7.14.1 Commercially available solutions may be used in place of laboratory preparations provided the solutions have been certified as equivalent.

7.14.2 Alternate volumes of solutions may be prepared provided the final solution concentration is equivalent.

⁷ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

7.15 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water as defined by Type III of Specification D 1193.

7.16 *Titration Solvent*—In a brown reagent bottle, add 30 mL of water to 1 L of isopropyl alcohol and mix thoroughly. Add 1 L each of toluene and chloroform and mix thoroughly.

7.17 *Toluene*, reagent grade. (**Warning**—Extremely flammable.)

7.18 *2,4,6-Trimethyl Pyridine (γ -Collidine)*, $(\text{CH}_3)_3\text{C}_5\text{H}_2\text{N}$, (**Warning**—Toxic), (mol weight 121.18) conforming to the following requirements:

Boiling range	168 to 170°C
Refractive index N_D^{20}	1.4982 \pm 0.0005
Color	colorless

7.18.1 Store the reagent over activated alumina and keep it in a brown glass bottle.

8. Preparation of Electrode System

8.1 *Maintenance of Electrodes*—Clean the glass electrode (Note 5) at frequent intervals (not less than once every week during continual use) by immersing in cold chromic acid cleaning solution (**Warning**—Causes severe burns. Recognized carcinogen. Strong oxidizer). Drain the calomel electrode at least once each week and refill with fresh LiCl electrolyte as far as the filling hole. Make sure that crystallized LiCl is present in the solution. Maintain the electrolyte level in the reference electrode above that of the liquid in the titration beaker at all times. When not in use, immerse the lower halves of the electrodes in either water (glass) or the LiCl in isopropyl alcohol electrolyte (reference). Do not allow them to remain immersed in titration solvent for any appreciable period of time between titrations. While the electrodes are not extremely fragile, handle them carefully at all times.

NOTE 5—Cleaning the electrodes thoroughly, keeping the ground-glass joint free of foreign materials, and regular testing of the electrodes are important in obtaining repeatable potentials, since contamination can introduce uncertain, erratic, and unnoticeable liquid contact potentials.⁸ While this is of secondary importance when end points are chosen from inflection points in the titration curve, it is quite serious when end points are chosen at arbitrarily-fixed cell potentials (that is, the nonaqueous acidic buffer potential).

8.2 *Preparation of Electrodes*, Before and after using, blot-dry the glass electrode thoroughly with a clean cloth or a soft absorbent tissue and rinse with water. Wipe the reference electrode with a cloth or tissue, carefully remove the ground-glass sieve, and thoroughly wipe both ground surfaces. Replace the sleeve loosely and allow a few drops of electrolyte to drain through to flush the ground-glass joint (Note 5). Wet the ground surfaces thoroughly with electrolyte, set the sleeve firmly in place, and rinse the electrode with water. Prior to each titration, soak the prepared electrodes in water for at least 5 min immediately before use and touch the tips of the electrodes with a dry cloth or tissue to remove the excess water.

⁸ For a detailed discussion of the need for care in preparation of the electrodes, see Lykken, L., Porter, P., Ruliffson, H. D., and Tuemmler, F. D., "Potentiometric Determination of Acidity in Highly Colored Oils," *Industrial and Engineering Chemistry*, Analytical Edition, IENAA, Vol 16. 1944, pp. 219–234.

8.3 *Testing of Electrodes*—Test the meter-electrode combination (Note 6) when first put into use or when new electrodes are installed and retest at intervals thereafter by dipping the electrodes into a well-stirred mixture of 100 mL of the titration solvent and 1.0 to 1.5 mL of 0.1 M alcoholic KOH solution. For the meter-electrode combination to be suitable for use, the potential between the electrodes must change by more than 0.480 V from the potential between the same electrodes when dipped in the nonaqueous acidic buffer solution.

NOTE 6—Considerably more sensitive electrodes are now available that will show a potential change of at least 0.590 V under these conditions, and their use is recommended.

9. Standardization of Apparatus

9.1 *Determination of Meter Readings for the Nonaqueous Buffer Solution Corresponding to Base End Point*—To ensure comparable selection of end points with the meter described in A1.1.1, determine daily for each electrode pair, the meter reading obtained with the freshly prepared nonaqueous acidic buffer solution to be used for the determination of base numbers, and with the freshly prepared nonaqueous basic buffer solution to be used for the determination of strong base numbers.

NOTE 7—The response of different glass electrodes to hydrogen ion activity is not the same. Therefore, it is necessary to establish regularly for each electrode system the meter readings corresponding to the acidic and basic buffer solutions arbitrarily selected to represent the end point.

9.2 Prepare the electrodes as described in 8.2, immerse them in the appropriate nonaqueous buffer solution, and stir for 5 min, maintaining the temperature of the buffer solution at a temperature within 2°C of that at which the titrations are to be made. Read the cell voltage. The reading so obtained in the acidic buffer solution is taken as the end point for the base number if an inflection is not observed as specified in 12.1, and the reading obtained in the basic buffer solution is taken as the end point for the strong base number.

10. Preparation of Sample of Used Oil

10.1 Strict observation of the sampling procedure is necessary, since the sediment itself is acidic or basic or has absorbed acidic or basic material from the sample. Failure to obtain a representative sample negates a meaningful value obtained.

NOTE 8—As used oils can change appreciably in storage, samples should be tested as soon as possible after removal from the lubricating system; and the dates of sampling and testing shall be noted.

10.2 Heat the sample (Note 9) of used oil to $60 \pm 5^\circ\text{C}$ in the original container and agitate the sample until all of the sediment is homogeneously suspended in the oil. If the original container is a can, or if it is glass and more than three-fourths full, transfer the entire sample to a clear-glass bottle having a capacity at least one-third greater than the volume of the sample. Transfer all traces of sediment from the original container to the bottle by vigorous agitation of portions of the sample in the original container.

10.3 After complete suspension of all sediment, strain the sample or a convenient aliquot through a 150- μm (100-mesh) screen for the removal of large contaminated particles.

NOTE 9—When samples are visibly free of sediment, the heating

procedures described can be omitted.

11. Procedure for Base Number and Strong Base Number

11.1 Calculate the quantity of sample required for its expected base number as follows:

$$A = 7/E \quad (1)$$

where:

A = approximate mass of sample, g and

E = expected base number.

11.1.1 Take a maximum of 5 g and a minimum of 0.1 g for analysis. The precision of weighing is as follows:

Size of Sample, g	Precision of Weighing g
1–5	0.005
0.1–1	0.002

11.2 Into a 250-mL titration beaker or a suitable titration vessel, introduce a weighed quantity of sample as prescribed in 11.1.1 and add 125 mL of titration solvent (Note 10). Prepare the electrodes as directed in 8.2. Place the beaker or titration vessel on the titration stand and adjust its position so that the electrodes are about half immersed. Start the stirrer, and stir throughout the determination at a rate sufficient to produce vigorous agitation without spattering and without stirring air into the solution.

NOTE 10—Some automatic titrators do not accept a beaker size that contains 125 mL of titration solvent. In such cases, a lesser amount of solvent in the range from 75 to 100 mL is acceptable.

11.3 Select and fill a suitable buret with the 0.1 M alcoholic HCl solution and place the buret in position on the titration assembly, taking care that the tip is immersed about 25 mm in the liquid in titration vessel. Record the initial buret and meter (cell potential) readings.

11.4 *Titration*—The reaction of the hydrochloric acid with the basic components is very slow with most titrations for base number. As a result, these titrations are not at equilibrium. Because of this, the titration conditions are tightly specified and must be strictly adhered to in order to achieve the precision as stated.

NOTE 11—See Appendix X1 for techniques for reducing the titration time of a sample. Pre-dosing techniques have been found to provide satisfactory results although the precision using these techniques has not been determined.

11.4.1 Whether the titration is carried out manually or automatically, the following procedure of *fixed increment, fixed time* addition of titrant must be followed. Add 0.1 M HCl in increments of 0.100 mL throughout the titration with a 90-s pause between each incremental addition. Take millivolt readings at the end of each 90 s interval. Continue as above until a potential is reached which is 100 mV past the meter reading corresponding to that found for the standard acidic buffer solution (acidic buffer potential). If the volume of titrant required to reach this potential (100 mV past the acidic buffer potential) is greater than 4.0 mL, reduce the sample size by one-half and repeat the titration.

11.4.1.1 The meter readings of potential difference are plotted manually or automatically against the respective volumes of titrant, and the end point taken as described in 12.1.

11.4.2 On completion of the titration, remove the titration vessel and rinse the electrodes and buret tip with the titration solvent, then with water, then again with titration solvent. (Soak electrodes in distilled water for at least 5 min before using for another titration.) Store the glass electrode in deionized or distilled water and the reference electrode in a saturated solution of LiCl in isopropyl alcohol when not in use (see 8.1).

11.4.3 *Blanks*—For each set of samples, make a blank titration of the same volume of titration solvent used for the sample. For the base number blank, add 0.1 M alcoholic HCl solution in 0.05-mL increments, waiting 90 s between each addition, until a potential which is 100 mV past the buffer potential (see 11.4) is reached. For the strong base number blank, add titrant under the same conditions until the potential corresponding to the basic buffer solution is reached.


12. Calculation

12.1 If an inflection (see Note 12) occurs in the potential region between the acidic buffer potential (see 11.4) and a point 100 mV past this potential, mark this inflection as the end point. If no inflection occurs in the above mentioned potential region, mark as the end point the point on the curve that corresponds to the acidic buffer potential. See Fig. 2 for examples of end points.

NOTE 12—An inflection point is generally recognizable by inspection whenever at least five successive cell potential changes, Δ, caused by the addition of the corresponding five increments of titrant, exhibit a maximum as illustrated by an example in the following table.

Titrant, mL	Δ, mV
1.8	8.3
1.9	10.7
2.0	11.3
2.1	10.0
2.2	7.9

The Δ at the maximum should be at least 5 mV, and the difference in Δ between the maximum and both the first and last Δ should be at least 2 mV.



- (A) Titration curve has no inflections. Take end point at the buffer potential.
- (B) Titration curve has inflection within prescribed window. Take inflection as the end point.
- (C) Titration curve has inflection prior to buffer potential but not in prescribed window. Take end point at buffer potential.

FIG. 2 Example Titration Curves to Illustrate Selection of End Points

12.2 Calculate the base number and strong base number as follows:

$$\text{Base number, mg KOH/g} = [(A-B) \times M \times 56.1]/W \quad (2)$$

$$\text{Strong base number, mg KOH/g} = [(CM + Dm) \times 56.1]/W \quad (3)$$

where:

A = alcoholic HCl solution, mL, used to titrate the sample to the end point (nonaqueous acidic buffer or inflection—see 12.1),

B = alcoholic HCl, mL, used to titrate the solvent blank to the same potential at which the sample end point occurs,

M = molarity of the alcoholic HCl solution,

W = sample, g,

C = alcoholic HCl solution, mL, used to titrate the sample to an end point that occurs at a meter reading corresponding to the nonaqueous basic buffer (see 7.2)

D = alcoholic KOH solution, mL, used to titrate the solvent blank to the potential corresponding to C, and

m = molarity of the alcoholic KOH solution.

13. Report

13.1 Report the results as base number, and strong base number, Test Method D 4739.

14. Quality Control Checks

14.1 confirm the performance of the test procedure by analyzing a quality control (QC) sample that is, if possible, representative of the samples typically analyzed.

14.2 Prior to monitoring the measurement process, the user of the method needs to determine the average value and control limits of the QC sample.

14.3 Record the QC results and analyze by control charts or other statistically equivalent technique to ascertain the statistical control status of the total testing process. Any out of control data should trigger investigation for root cause(s). The results of this investigation may, but not necessarily, result in instrument recalibration.

14.4 The frequency of QC testing is dependent on the criticality of the quality being measured, the demonstrated stability of the testing process, and customer requirement. Generally, a QC sample should be analyzed each testing day. The QC frequency should be increased if a large number of samples are routinely analyzed. However, when it is demonstrated that the testing is under statistical control, the QC testing frequency may be reduced.

15. Precision and Bias ⁹

15.1 *Precision*—The precision of this test method as determined by statistical examination of results on nine samples of new and used oils run in duplicate by twelve different laboratories is as follows:

15.1.1 *Base Number:*

15.1.1.1 *Repeatability*—The difference between two test results, obtained by the same operator with the same apparatus

⁹ Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR: D02-1217.

under constant operating conditions on identical test material, would, in the long run, in the normal and correct operation of the test method, exceed the following value only in one case in twenty:

$$10.4 \% \text{ of the mean of the two test results} \quad (4)$$

15.1.1.2 *Reproducibility*—The difference between two single and independent results obtained by different operators working in different laboratories on identical test material, would, in the long run, in the normal and correct operation of the test method, exceed the following value only in one case in twenty:

$$21.1 \% \text{ of the mean of the two test results} \quad (5)$$

NOTE 13—The range of base number values for which the precision values were established was 0.5 to 70.

15.1.2 *Strong Base Number*—Precision data have not been developed for strong base number because of its rare occurrence.

15.2 *Bias*—A statement of bias is not applicable since a standard reference material for this property is not available.

16. Keywords

16.1 base number; lubricants; petroleum products; potentiometric titration

ANNEX

(Mandatory Information)

A1. APPARATUS

A1.1 *Apparatus for Manual Titration*—Shall consist of the following:

A1.1.1 *Meter*—A voltmeter or potentiometer that will operate with an accuracy of ± 0.005 V and a sensitivity of ± 0.002 V, over a range of at least ± 0.5 V, when the meter is used with the electrodes specified in A1.1.2 and A1.1.3, and when the resistance between the electrodes falls within the range from 0.2 to 20 M Ω . The meter shall be protected from stray electrostatic fields so that no permanent change in the meter readings over the entire operating range is produced by touching with a grounded lead (Note A1.1), any part of the exposed surface of the glass electrode, the glass electrode lead, the titration stand, or the meter. A desirable apparatus may consist of a continuous-reading electronic voltmeter with specified range, accuracy, and sensitivity, that is designed to operate on an input of less than 5×10^{-12} A when an electrode system having 1000 M Ω resistance is connected across the meter terminals, and that is provided with a satisfactory terminal to connect the shielded connection wire from the glass electrode to the meter without interferences from the presence of external electrostatic field.

NOTE A1.1—*Grounded or connected to the ground* means connected through a resistance of not more than 100 Ω to a standard ground potential such as that of a water-service pipe.

A1.1.2 *Glass Electrode*—A pencil-type glass electrode (C, Fig. 1) 125 to 180 mm in length and 8 to 14 mm in diameter. The body of the electrode shall be made of a chemically resistant glass tube with a wall thickness of 1 to 3 mm. The end dipping into the solution shall be closed with a hemisphere of Corning 015 glass or equivalent sealed on to the electrode tube, and the radius of this hemisphere shall be about 7 mm. The thickness of the glass in the hemisphere shall be great enough so that the resistance of the hemisphere is 100 to 1000 M Ω at 25°C. The electrode shall contain a reproducible, permanently sealed liquid cell for making electrical connection with the

inner surface of the hemisphere. The entire electrical connection from the sealed contact cell to the meter terminal shall be surrounded by an electrical shield that will prevent electrostatic interferences when the shield is grounded. The shield shall be insulated from the electrical connection by insulating material of the highest quality, such as rubber and glass, so that the resistance between the shield and the entire length of the electrical connection is greater than 50 000 M Ω .

A1.1.3 *Calomel Electrode*—A pencil-type calomel electrode (B, Fig. 1) 125 to 180 mm in length and 8 to 14 mm in diameter. This electrode shall be made of glass and shall be provided with an external, removable glass sleeve on the sealed end that is dipped into the titration solution. The glass sleeve shall be 8 to 25 mm in length, shall be slightly tapered, and shall be ground to fit the electrode so that the sealed end of the electrode protrudes 2 to 20 mm beyond the sleeve. The ground surface shall be continuous and free of smooth spots. At a point midway between the extremities of the ground surface, the electrode tube shall be pierced by a hole or holes 1 mm in diameter. The electrode shall contain the necessary mercury, calomel and electrical connection to mercury, (calomel), or silver-silver chloride (Ag/AgCl), all arranged in a permanent manner. The electrode shall be filled almost to capacity with saturated LiCl in isopropyl alcohol electrolyte and shall be equipped with a stoppered port through which the electrolyte may be replenished. When suspended in the air and with the sleeve in place, the electrode shall not leak electrolyte at a rate greater than one drop in 10 min.

A1.1.4 *Stirrer*—A variable-speed mechanical stirrer of any suitable type, equipped with a glass, propeller-type-stirring paddle (D, Fig. 1). A propeller with blades 6 mm in radius and set at a pitch of 30 to 45° is satisfactory. A magnetic stirrer is also satisfactory. If electrical stirring apparatus is used, it must be grounded so that connecting or disconnecting the power to

the motor will not produce a permanent change in meter reading during the courses of titration.

A1.1.5 *Buret*—A 5-mL buret (E, Fig. 1) graduated in 0.01-mL divisions and calibrated with an accuracy of ± 0.005 mL. The buret shall have a glass stopcock and shall have a tip that extends 100 to 130 mm beyond the stopcock.

A1.1.6 *Titration Beaker*—A 250-mL beaker made of borosilicate glass, or other suitable titration beaker, (A, Fig. 1).

A1.1.7 *Titration Stand*—A suitable stand to support the electrodes, stirrer, and buret in the position shown in Fig. 1. An arrangement that allows the removal of the beaker without disturbing the electrodes, buret, and stirrer is desirable.

A1.2 Automatic titration system shall be generally in

accordance with A1.1 and provide the following technical performance characteristics of features:

A1.2.1 The addition of titrant must be automatically controlled to dispense discontinuously 0.100 ± 0.001 -mL increments of titrant with a waiting period of 90 s between increments.

A1.2.2 Interchangeable precision motor-driven burets with volume dispensing accuracy of ± 0.001 mL.

A1.2.3 A record of the complete course of a titration by continuously printing out the potential or change in potential with the addition of each increment of titrant, versus volume of titrant added.

APPENDIX

(Nonmandatory Information)

X1. REDUCING TITRATION TIME

X1.1 A long, equilibration period of 90 s/increment was selected for the base number titration because the titration reaction and electrode equilibration are generally slow. This, of course, can lead to long titration time/sample, with a maximum time of 1 h based on a maximum volume of titrant of 4 mL and a rate of titrant addition of 0.1 mL/90 s. It is possible to substantially reduce the titration time by predosing with rapid addition of titrant until a potential within 25 mV of the buffer potential is reached, then allowing 90 s for equilibration and completing the titration under normal conditions. This procedure is not expected to have an adverse affect on the precision of this test method; however, the precision under these conditions has not been determined.

X1.2 There are many cases where the optimum in precision in the base number is not required, and in these cases the titration time can be shortened by taking a smaller sample. For example, for a base number of 2 and using a sample size calculated from the equation in 11.1, a total titration time of 24 min would be required. By taking only one half of the prescribed sample size the titration time would be reduced to 12 min. The affect of halving the sample size on the precision of this test method has not been determined but it would be expected to be small.

SUMMARY OF CHANGES

Subcommittee D02.06 has identified the location of selected changes to this standard since the last issue (D 4739-96) that may impact the use of this standard.

- (1) Placed relation to former base number portion in Test Method D 664 in the scope.
- (2) Added references to ASTM practices for manual and automatic sampling.
- (3) Updated title of reference IP 139.
- (4) Placed permission for commercial solutions and variations of reagent volumes in 7.14.1 and 7.14.2.

- (5) Clarified definitions and corrected chemical name for isopropyl alcohol.
- (6) Removed reference to use of Calomel electrodes.
- (7) Inserted a new warning statement concerning possible peroxide formation in propanol-2-ol.
- (8) Added Section 14 on Quality Control Checks.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).