



# Standard Test Method for Determination of pHe of Ethanol, Denatured Fuel Ethanol, and Fuel Ethanol (Ed75-Ed85)<sup>1</sup>

This standard is issued under the fixed designation D 6423; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This test method covers a procedure to determine a measure of the acid strength of high ethanol content fuels. These include ethanol, denatured fuel ethanol, and fuel ethanol (Ed75-Ed85). The test method is applicable to fuels containing nominally 70 volume % ethanol, or higher, as described in Specifications D 4806 and D 5798.

1.2 Acid strength as measured in this test method is defined as pHe. A pHe value for alcohol solutions is not directly comparable to pH values of water solutions.

1.3 The value of pHe will depend somewhat on the fuel blend, the stirring rate, and the time the electrode is in the fuel.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

## 2. Referenced Documents

### 2.1 ASTM Standards:

D 4806 Specification for Denatured Fuel Ethanol for Blending with Gasolines for Use as Automotive Spark-Ignition Engine Fuel<sup>2</sup>

D 5798 Specification for Fuel Ethanol (Ed75-Ed85) for Automotive Spark-Ignition Engines<sup>2</sup>

## 3. Terminology

### 3.1 Definitions:

3.1.1 *ethanol, n*—ethyl alcohol, the chemical compound, C<sub>2</sub>H<sub>5</sub>OH.

3.1.2 *denaturants, n*—natural gasoline, gasoline components, unleaded gasoline, or toxic or noxious materials added to fuel ethanol to make it unsuitable for beverage use but not unsuitable for automotive fuel use.

3.1.3 *denatured fuel ethanol, n*—fuel ethanol made unfit for beverage use by the addition of denaturants.

3.1.4 *fuel ethanol (Ed75-Ed85), n*—a blend of ethanol and hydrocarbons of which the ethanol portion is nominally 75 to 85 volume % denatured ethanol.

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D-2 on Petroleum Products and Lubricants and is the direct responsibility of Subcommittee D02.A on Gasoline and Oxygenated Fuels.

Current edition approved June 10, 1999. Published August 1999.

<sup>2</sup> *Annual Book of ASTM Standards*, Vol 05.03.

### 3.2 Definitions of Terms Specific to This Standard:

3.2.1 *pHe, n*—a measure of the acid strength of alcohol fuels defined by this apparatus and procedure.

## 4. Summary of Test Method

4.1 The sample is analyzed at room temperature using a specified electrode system and a pH meter with a sufficiently high impedance and recommended for use with ion specific electrodes.

4.2 Readings are taken at exactly 30 s because the meter reading will drift throughout the analysis due to solvent effects on the electrode.

4.3 The electrode is soaked in water-based pH 7 buffer between readings to prepare it for the next sample. This rehydrates the glass electrode, a necessary step to preserve the electrode's response characteristics.

## 5. Significance and Use

5.1 The acid strength, as measured by pHe, is a good predictor of the corrosion potential of ethanol fuels. It is preferable to total acidity because total acidity does not measure acid strength; overestimates the contribution of weak acids, such as carbonic acid; and may underestimate the corrosion potential of low concentrations of strong acids, such as sulfuric acid.

## 6. Apparatus

6.1 *pH meter*—All types of commercially available pH meters with a sufficiently high impedance and recommended for use with ion specific electrodes are acceptable for this test method. Temperature compensation and readability to 0.01 pH unit are recommended.

6.2 *Electrode*—ORION Ross Sure-Flow combination electrode<sup>3</sup> with a glass body shall be used. Because the measurement is (of necessity) not made at equilibrium, it is essential that this exact electrode be used to ensure the reproducibility of results. Other electrodes (even those of similar design) will likely give different results under some or all conditions due to the use of a different size or type of glass membrane for the pH electrode, a different type of salt bridge junction, or other small

<sup>3</sup> The specified electrode (ORION Cat. No. 8172BN) is available from many laboratory supply companies or from the manufacturer: Orion Research Incorporated, Laboratory Products Group, The Schrafft Center, 529 Main Street, Boston, MA 02129.

differences, which may affect their nonequilibrium response.

6.3 *Temperature Compensator*—The thermocompensator is a temperature-sensitive resistance element immersed in the sample with the electrodes. The thermocompensator automatically corrects for the change in slope of the glass electrode response (with change in temperature) but does not correct for actual changes in sample pH with temperature. Because temperature compensation corrects only for changes in pH electrode response with temperature, the fuel sample must be at  $22 \pm 2^\circ\text{C}$ .

6.4 *Beakers*, borosilicate glass, 100 mL.

6.5 *Magnetic Stirrer*—Any laboratory magnetic stirrer can be used, along with a TFE-fluorocarbon-coated stirring bar approximately 19 to 25-mm long.

6.6 *Timer*, capable of measuring seconds.

## 7. Reagents and Materials

7.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.<sup>4</sup> Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

7.2 *Buffer Solutions*—Commercially available, prepared buffer solutions (pH=4.00 and pH=7.00) are acceptable for the standardization.

7.3 *1 M Hydrochloric Acid (HCl)*—Mix one volume of concentrated (12 M) HCl with eleven volumes of distilled water.

7.4 *3 M Potassium Chloride (KCl)*—The electrode is shipped with this filling solution already prepared.

7.5 *1 M Sodium Hydroxide (NaOH)*—Dissolve 4 g of NaOH pellets in 100 mL of distilled water.

7.6 *1 M Sulfuric Acid (H<sub>2</sub>SO<sub>4</sub>)*—Mix one volume of concentrated (18 M) H<sub>2</sub>SO<sub>4</sub> with 17 volumes of distilled water.

## 8. Sample Containers

8.1 Care must be used in sample container selection to avoid reactions that change the pHe of the sample. Do not use steel, plated steel, or epoxy-coated steel containers or PVC-coated glass bottles for samples. Avoid closures with a paper seal. Fluorinated high-density polyethylene and spun aluminum sample containers with polyethylene closures are acceptable. Glass bottles without a PVC coating and utilizing TFE-fluorocarbon or polyethylene cap seals are also acceptable.

8.2 Rinse the container with a portion of the fuel to be sampled before taking the sample.

## 9. Standardization of Assembly

9.1 Turn on the pH meter and allow it to warm up thoroughly in accordance with the manufacturer's instructions.

<sup>4</sup> *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, D.C. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

Note the temperature of the sample to be tested. If temperature compensation is to be manual, adjust the temperature dial of the meter to correspond to the temperature of the sample to be tested and allow time for all buffers, solutions, and the electrode to equilibrate thermally.

9.2 Clean/rehydrate the electrode after every ten samples and new electrodes before first use in alcohol solutions by alternately soaking several times in 1 M NaOH solution and 1 M H<sub>2</sub>SO<sub>4</sub> (or 1 M HCl) for about 30 s each. Remove the electrode and rinse it well with distilled water into a waste container.

9.3 Calibrate the pH meter to pH=7.00 with a water-based pH=7.00 buffer solution. Remove the electrode and rinse it with distilled water into a waste container.

9.4 Calibrate the pH meter to pH=4.00 with a water-based pH=4.00 buffer solution, using the *slope* adjustment. The slope must be in the 95 to 100 % range, or the electrode will have to be cleaned or replaced. Remove the electrode and rinse it with distilled water into a waste container. Return the electrode to the pH=7.00 buffer.

## 10. Procedure

10.1 Fill the electrode as needed with 3M KCl solution, which is supplied with the electrode.

10.2 Standardize the assembly with two reference buffer solutions, as described in 9.3 to 9.4.

10.3 Start with a new aliquot (fresh portion) of ethanol sample for each measurement. (Poor repeatability can be partly due to a poorly buffered sample. High-purity solutions are poorly buffered because they do not contain compounds that tend to stabilize pH. Since some alcohols meet this criteria, it is important that a fresh sample be used for each measurement and that electrodes be in top condition.)

10.4 Place about 50 mL of sample in a 100-mL beaker containing a magnetic stirring bar and either a thermometer (for meters with manual temperature compensation) or an ATC probe (for meters with automatic temperature compensation), and ensure the sample is at  $22 \pm 2^\circ\text{C}$ .

10.5 Remove the electrode from the pH=7.00 buffer, and rinse the electrode with distilled water into a waste container.

10.6 Blot the electrode to remove excess water.

10.7 Insert the electrode into the room-temperature stirred ethanol sample, start the timer, and measure the pHe at exactly  $30 \pm 1$  s. Stirring must be rapid enough to cause a small to moderate vortex about 6 to 8 mm-deep. Note that the reading will still be slowly changing after 30 s due to dehydration of the glass pH electrode, which causes a shift in its output. If the reading is not rising at the 30-s point for low pHe fuels (pHe<5), or gives poor repeatability in some solutions (those with low buffering capacity), the electrode probably needs to be cleaned or replaced.

10.8 Remove the electrode from the sample and rinse it with distilled water into a waste container. Return the electrode to the pH=7.00 buffer to rehydrate the glass pH electrode.

10.9 Soak the electrode in the buffer until the reading falls below 7.05, but soak for at least 20 s. Recalibrate to pH=7.00 if the pH is still above 7.05 after 5 min.

10.10 Clean/rehydrate the electrode as described in 9.2 after ten alcohol samples have been measured. The electrode may

give poor repeatability or high results (in alcohols) without periodic pretreatment by soaking in acid and base before use, probably because the glass electrode is contaminated or not fully hydrated.

10.11 Start with step 10.3 for the next sample.

## 11. Precision and Bias

11.1 *Precision*—The precision of this test method as determined by statistical examination of the interlaboratory test results is as follows:

11.1.1 *Repeatability*—The difference between successive test results, obtained by the same operator using the same apparatus under constant operating conditions on identical test material, would in the long run, in the normal and correct

operation of this test method, exceed 0.29 pHe in only one case in twenty.

11.1.2 *Reproducibility*—The difference between two single and independent test results, obtained by different operators working in different laboratories on identical test material, would in the long run, in the normal and correct operation of this test method, exceed 0.52 pHe in only one case in twenty.

11.2 *Bias*—Since there is no material having an accepted reference value for pHe, bias cannot be determined.

## 12. Keywords

12.1 acid strength; automotive spark ignition engine fuel; corrosion potential; denatured fuel ethanol; ethanol; fuel ethanol; pHe

*The American Society for Testing and Materials takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.*

*This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.*

*This standard is copyrighted by ASTM, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).*