

Standard Test Method for Assay of Di-*tert*-Butyl Peroxide Using Gas Chromatography¹

This standard is issued under the fixed designation E 475; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the assay of di- *tert*-butyl peroxide using gas chromatography. It is applicable to commercial di-*tert*-butyl peroxide which may contain small amounts of isobutylene, *tert*-butanol, *tert*-butyl hydroperoxide, triisobutylenes, and water as impurities.

1.2 Review the current appropriate Material Safety Data Sheets (MSDS) for detailed information concerning toxicity, first aid procedures, and safety precautions.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* Specific hazards statements are given in Section 9.

2. Referenced Documents

2.1 ASTM Standards:

- E 180 Practice for Determining the Precision of ASTM Methods for Analysis and Testing of Industrial Chemicals²
- E 203 Test Method for Water Using Karl Fischer Reagent²

3. Terminology

3.1 Definitions:

3.1.1 *active oxygen*—the oxidizing power present in organic peroxides expressed as oxygen (equivalent weight = 8.00).

4. Summary of Test Method

4.1 A sample is diluted in dodecane and injected into a gas chromatograph containing a wide-bore capillary column with a nonpolar stationary phase. A temperature program is used to separate di- *tert*-butyl peroxide from impurities using helium as a carrier gas. The flame ionization detector response, proportional to component concentration, is recorded and the peak areas are measured. The concentration of di-*tert*-butyl peroxide is determined by area normalization after correcting for the area of the solvent peak. The assay of di- *tert*-butyl peroxide is corrected for water, which is determined using a separate Karl Fischer titration since water is not detected using flame

ionization detection. The area percent of di-*tert*-butyl peroxide is assumed to be equal to weight percent.

5. Significance and Use

5.1 Di-*tert*-butyl peroxide is widely used as a catalyst and reaction initiator. Knowledge of the peroxide content is important in such applications. This test method provides a procedure for determining the active peroxide content of commercial di- *tert*-butyl peroxide.

6. Interferences

6.1 Interferences will be encountered if other components are present in the sample that have the same retention time as di- *tert*-butyl peroxide.

7. Apparatus

7.1 *Instrumentation*— Gas chromatograph, capable of column oven temperature programming from 40 to 160°C at a rate of 10°C/min and from 160 to 240°C at a rate of 35°C/min.

7.2 *Injection System*— Glass-lined sample injection port, maintained at 150°C.

7.3 *Sample Introduction*—Microliter syringes or automatic syringe injectors, capable of delivering 0.5 μ L of a liquid sample, have been used successfully.

7.4 *Detection*—Flame ionization detector, maintained at 250°C.

7.5 *Data Acquisition System*—Electronic data acquisition and area integration capabilities are recommended.

7.6 *Chromatographic Column*—30 m wide-bore capillary column (0.53 mm inside diameter) with a 5.0 μ m dimethylpolysiloxane stationary phase.³

8. Reagents and Materials

8.1 *Carrier Gas (Helium)*, chromatographic grade.

8.2 *Hydrogen*, chromatographic grade.

8.3 *Compressed Air*, oil-free.

8.4 *Dodecane*, for use as a sample diluent, 99 % or greater chromatographic purity.

9. Hazards

9.1 Organic peroxides are strong oxidizing agents and present potential fire and explosion hazards. While di- *tert*-butyl peroxide is one of the more stable peroxides, contact with

¹ This test method is under the jurisdiction of ASTM Committee E15 on Industrial and Specialty Chemicals and is the direct responsibility of Subcommittee E15.02 on Product Standards.

Current edition approved Oct. 10, 2000. Published November 2000. Originally published as E 475 – 73. Last previous edition E 475 – 95 (1999) ^{ϵ 1}.

² *Annual Book of ASTM Standards*, Vol 15.05.

³ A J & W Scientific DB-1, available from J & W Scientific, 91 Blue Ravine Rd., Folsom, CA 95630-4714, was utilized to develop this test method.

reducing agents and sources of heat, sparks, or open flames must be avoided. Organic peroxides in general are irritating to the skin, eyes, and mucous membranes. Avoid bodily contact and handle only in a well-ventilated area.

10. Procedure

10.1 Install the column in the chromatograph. The operating conditions required to give the desired separation are given in Table 1. Set the helium flow to 10.0 mL/min and the oven temperature to 40°C and allow sufficient time for the instrument to equilibrate as indicated by a stable baseline.

10.2 Prepare test samples by transferring approximately 0.2 g of di-*tert*-butyl peroxide into a flask or vial and add approximately 10 mL of dodecane. Mix well.

10.3 Inject 0.5 µL of the sample solution into the chromatograph. A typical chromatogram is shown in Fig. 1 and Fig. 2.

10.4 Determine the water content of the sample using a Karl Fischer titration in accordance with Test Method E 203.

NOTE 1—Water can be determined in samples of di-*tert*-butyl peroxide using Test Method E 203 without pretreatment with SO₂ if the sample temperature is maintained below 20°C during the titration. If the sample temperature cannot be maintained below 20°C, the sample must be pretreated by bubbling SO₂ through the sample for several minutes prior to titrating.

11. Calculation

11.1 Measure the peak areas and calculate the total area by summing the individual areas for all peaks except for that of the solvent peak, as follows:

$$A_t = A_1 + A_2 + A_3 + \dots + A_n \quad (1)$$

where:

A_t = total area (excluding the solvent peak), and

$A_{1..n}$ = area of individual component peaks.

11.2 Calculate the area percent of di-*tert*-butyl peroxide as follows:

$$\text{di-tert-butyl peroxide, area \%} = \frac{A_D(100 - W)}{A_t} \quad (2)$$

where:

A_D = area of the di-*tert*-butyl peroxide peak,

A_t = sum total of all component peaks (except solvent), and

W = weight percent of water in the sample, as determined in 10.4.

The area percent of di-*tert*-butyl peroxide is assumed to be equal to the weight percent.

11.3 If desired, calculate the weight percent of active oxygen due to di-*tert*-butyl peroxide as follows:

$$\text{active oxygen, weight \%} = \frac{C \times 8.000}{73.11} \quad (3)$$

where:

C = di-*tert*-butyl peroxide, weight %.

12. Report

12.1 Report the purity of the di-*tert*-butyl peroxide to the nearest 0.01 %.

13. Precision and Bias

13.1 *Precision:*

13.1.1 *Repeatability (Single Analyst)*—The coefficient of variation for a single determination has been estimated to be 0.05 % relative at 18 degrees of freedom (DF). The 95 % limit for the difference between two such runs is 0.13 % relative.

13.1.2 *Laboratory Precision (Within-Laboratory, Between-Days Variability, Formerly Called Repeatability)*—The coefficient of variation of results (each the average of duplicates) obtained by the same analyst on different days has been estimated to be 0.04 % relative at 9 DF. The 95 % limit for the difference between two such averages is 0.10 % relative.

13.1.3 *Reproducibility (Multilaboratory)*—The coefficient of variation of results (each the average of duplicates) obtained by analysts in different laboratories has been estimated to be 0.24 % relative at 8 DF. The 95 % limit for the difference between two such averages is 0.67 % relative.

NOTE 2—The precision statements are based on an interlaboratory study⁴ performed in 1995 on one sample of di-*tert*-butyl-peroxide.

13.2 *Bias*—The bias of this test method has not been determined due to the unavailability of suitable reference materials.

14. Keywords

14.1 active oxygen; di-*tert*-butyl peroxide; gas chromatography; peroxides

⁴ Supporting data are available from ASTM Headquarters. Request RR:E15-1044.

TABLE 1 Operating Conditions Required to Give the Desired Separation

Detector	flame ionization (FID)
Temperatures, °C	
Injection port	150
Detector	250
Column temperature program	
Initial temperature, °C	40
Initial time, min	3
Program rate, °C/min	10
Temperature 2, °C	160
Time 2, min	2
Program rate, °C/min	35
Final temperature, °C	240
Final time, min	2
Carrier gas	helium
Flow rate, mL/min	10.0
Injection volume, µL	0.5




FIG. 1 Typical Chromatogram




FIG. 2 Typical Chromatogram (Expanded View)

The American Society for Testing and Materials takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).